

CERTIFIKAT O SKLADNOSTI CERTIFICATE OF COMFORMITY

naročnik applicant UTG VODNIK d.o.o.
Tolstojeva ulica 12, 1230 Domžale

lastnik owner UTG VODNIK d.o.o.
Tolstojeva ulica 12, 1230 Domžale

vzorec sample Medicinska maska za obraz
Medical face mask

proizvajalec manufacturer
UTG VODNIK

tip type Medicinska obrazna maska za enkratno uporabo, troslojna – 333383

Identifikacijska številka identification number	LOT številka LOT number	tip maske mask type
1-15	neznan / not known	II

preskusi tests	št. certifikata certificate No.	skladnost compliance
preskus učinkovitosti bakteriološke filtracije (BFE) bacterial filtration efficiency test (BFE)	589-3-20-1	✓
preskus na dihanje (zračna prepustnost) breathability (differential pressure test)	555-16-20-1	✓
preskus odpornosti na brizg tekočine splash resistance test	n/a	n/a
preskus čistosti na mikrobe (bio obremenitev) microbial cleanliness test (bioburden)	n/a	n/a
preskus biološke združljivosti (citotoksičnost) biocompatibility test (cytotoxicity)	10G014-2020	✓

n/a – se ne uporablja / not applicable

zaključek conclusion Vsi preskusi so pokazali, da maske **ustrezajo** zahtevam podanim v evropskem standardu EN 14683:2019+AC:2019, s tem da se rezultati nanašajo izključno na preskušane vzorce in izmerjene vrednosti v času meritev, ki ne zagotavljajo dolgotrajne stabilnosti ali razširjanje rezultatov na ostale vzorce ali serije vzorcev.
*All tests have shown the masks **meet** requirements given by European standard EN 14683:2019 +AC:2019, with results refer only to the tested item and to the measured values at the time of measurement, which carry no implication regarding the long term stability or disseminating results to other samples or series of samples.*

datum
date
07. 07. 2020

pripravil prepared by
Miha Zakotnik
Internally digitally signed



datum odobritve
date of approval

15.07.2020

odobril approved by
Primož Hafner
odgovorna oseba
responsible person



Ta certifikat v celoti nadomešča certifikat s številko 903-46-20-1, ki je bil izdan dne 07. 07. 2020.
This certificate fully replaces certificate with number 903-46-20-1, which was issued on 07. 07. 2020.

Ta dokument se sme objavljati ali posredovati le v celoti. Verodostojnost podpisa se lahko preveri v elektronski obliki dokumenta.
This document may be published or forwarded only in full. Signature validity can be verified in electronic version.